

HIV 1/2 STAT-PAK® DIPSTICK Assay

A rapid, qualitative screening assay for the detection of antibodies to HIV 1 & 2 in human serum, plasma or whole blood.

Convenient and Cost Effective

- Ideally suited for field and point-of-care testing
- Assay can be performed using card format
- Fingerstick or venous whole blood, serum or plasma
- Room temperature storage
- Long shelf life
- No special laboratory equipment required
- Economical shipping

Time and Labor Saving

- Ready-to-use reagents
- Simple procedure
- Rapid 15-minute results
- Results are easy to interpret

Minimal sample size: 5 µL

Reliable Results

- Built-in IgG procedural control
- Highly sensitive and specific:
Sensitivity: 99%-100%
Specificity: 100%*



Manufactured in the USA


 **CHEMBIO**
DIAGNOSTIC SYSTEMS, INC.
Rapid Tests for Earlier Treatment™

* See product insert for complete performance details

WHO Prequalified

Administering Test*


- 1** SEPARATE backing card from sheet.



- 2** REMOVE red liner from adhesive strip on back of DIPSTICK.




- 3** ATTACH DIPSTICK to backing card aligning arrows.




- 4** PLACE backing card and DIPSTICK assembly onto flat surface.




- 5** COLLECT sample using 5 μ L loop provided.



- 6** TOUCH sample to center of Sample (S) area on DIPSTICK.



- 7** ADD 3 drops of buffer to sample (S) area on DIPSTICK holding vial vertically.




- 8** READ test results between 15 and 20 minutes.




Interpretation of Results


+ REACTIVE




- NONREACTIVE



INVALID



INVALID



Ordering Information

* Before testing, please refer to the complete product insert provided in the test kit

Description	Contents	Catalog Number
HIV 1/2 STAT-PAK® DIPSTICK Assay WHO Prequalified	30 HIV 1/2 STAT-PAK® DIPSTICK Test Strips 1 Running Buffer Bottle (8 mL) 30 Disposable Sample Loops (5 μ L) 30 Backing Cards 1 Product Insert	HIV303 (60-9534-0)